


PRODUCT: Thawing Media

 LOT#: H0832
 EXPIRY DATE: 2018-12-13
 Year / Month / Date

 REF: 91182
 CODE: VT802


(01) 1 4582231 46077 6 (17) 181213 (10) H0832


 APPROVED BY: Kazumi Shibata
 Quality Assurance
 2018-09-13
CERTIFICATE OF ANALYSIS

Indication for Use: Thawing Media is used for thawing and recovery of vitrified oocytes and/or embryos.

Storage: Store between 2-8 °C.

Caution: Federal law restricts this device to sale by or on the order of a physician or practitioner trained in its use.

We certify that products manufactured by Kitazato Corporation have been approved in accordance with current Good Manufacturing Practice (cGMP). The quality of products was tested and evaluated by quality control procedures.

ASSAY	SPECIFICATION	RESULT			
		TS1	TS2	DS	WS
Color	Clear, Particle Free/ Rose Pink	Pass	Pass	Pass	Pass
pH	7.20-7.60	7.40	7.40	7.39	7.39
Osmolarity ¹	TS:1600-2000				
	DS:830-1020	1853	1855	944	285
	WS:240-300				
Endotoxin ²	< 0.25 EU/mL	<0.006	<0.006	<0.006	<0.006
Sterility ³	No Microbial Growth	Pass	Pass	Pass	Pass
MEA(One Cell) ⁴					
	-Control		100%		
	-This Lot		100%		

¹ Osmotic pressure is measured either by osmolarity; with freezing point in mOsm/L or by osmolality; with vapor pressure osmometry in mmol/kg.

² Bacteria Endotoxins Test, Photometric Techniques in accordance with the USP, EP and JP

³ Sterility Tests, Direct Inoculation in accordance with the USP, EP and JP

⁴ One cell mouse embryo testing is performed using fresh mouse embryos. Test results indicate the percentage of mouse embryos developing to fully expanded blastocysts after 96 hours in culture medium.