Gynotec

Certificate of Analysis

Product: SpermFilter® Gradient 80%

Batch number:	SPF80-701
Expiration Date: (DD/MM/YYYY)	30-09-2018

Analysis	Limits	Results
pH:	7,20 – 7,60	7,52
Osmolality:	290 – 310 mOsm/kg	305
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,0986
Sperm Survival:	≥ 80%	95%

Approved on: (DD/MM/YYYY)	19-04-2017
By:	M. Berbers
Signature:	

Quality Control Department, Gynotec B.V.

Ref. nr.: F052/2