



Certificate of Analysis

Product: SpermWash®

Batch number: SW-012
Expiration Date: 30-09-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,47
Osmolality:	270 – 290 mOsm/kg	281
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	100%

Approved on: 31-03-2020
(DD/MM/YYYY)

By: J. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'J. Berbers', is written over a light blue horizontal line.

Quality Control Department, Gynotec B.V.