



## Certificate of Analysis

### Product: SpermWash®

Batch number: SW-036  
Expiration Date: 30-04-2022  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,52
Osmolality:	270 – 290 mOsm/kg	280
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	100%

---

Approved on: 12-11-2020  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', with the name printed below it.

Quality Control Department, Gynotec B.V.