



Certificate of Analysis

Product: SpermWash®

Batch number: SW-124
Expiration Date: 31-12-2022
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,49
Osmolality:	270 – 290 mOsm/kg	283
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	100%

Approved on: 18-06-2021
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, light blue rectangular stamp.

Quality Control Department, Gynotec B.V.