



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-128
Expiration Date: 31-12-2022
(DD/MM/YYYY)

| Analysis | Limits | Results |
|--------------------------|-------------------|--------------|
| pH: | 7,30 – 7,60 | 7,46 |
| Osmolality: | 270 – 290 mOsm/kg | 283 |
| Sterility Test Bacteria: | NEGATIVE | PASSED |
| Sterility Test Fungi: | NEGATIVE | PASSED |
| Endotoxin: | < 0,25 EU/ml | < 0,05 EU/ml |
| MEA: | ≥ 80% | 100% |

Approved on: 22-07-2021
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp.

Quality Control Department, Gynotec B.V.