



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-155
Expiration Date: 30-06-2023
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,51
Osmolality:	270 – 290 mOsm/kg	279
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
(Sterile by filtration over 0,1µm filter)		
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
HSA concentration:	3,2 – 4,8 g/liter	4,597
MEA:	≥ 80%	100%
(1-cell to expanded blastocyst - 96 hours)		

Approved on: 12-01-2022
(DD/MM/YYYY)
By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.