



## Certificate of Analysis

### Product: SpermWash®

Batch number: SW-233  
Expiration Date: 31-01-2024  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,49
Osmolality:	270 – 290 mOsm/kg	279
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
(Sterile by filtration over 0,1µm filter)		
Endotoxin:	< 0,25 EU/ml	<0,05 EU/ml
HSA concentration:	3,2 – 4,8 g/liter	4,0
MEA:	≥ 80%	95%
(1-cell to expanded blastocyst - 96 hours)		

---

Approved on: 28-07-2022  
(DD/MM/YYYY)  
By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp or background.