



## Certificate of Analysis

### Product: SpermWash<sup>®</sup>

Batch number: SW-2420  
Expiration Date: 30-09-2025  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,39
Osmolality:	270 – 290 mOsm/kg	282
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	4,10
Gentamicin concentration:	5 – 15 mg/liter	8,44
MEA: (% blastocysts after 96 hours; 60 min. exposure time to test medium)	≥ 80%	100%
HSSA (motility compared with control after 24h exposure to test medium):	≥ 80%	100%

---

Approved on: 11-04-2024  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, circular stamp or watermark.