



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-2445
Expiration Date: 31-01-2026
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,39
Osmolality:	270 – 290 mOsm/kg	280
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	4,00
Gentamicin concentration:	5 – 15 mg/liter	9,06
MEA: (% blastocysts after 96 hours; 60 min. exposure time to test medium)	≥ 80%	95%
HSSA (motility compared with control after 24h exposure to test medium):	≥ 80%	94%

Approved on: 31-07-2024
(DD/MM/YYYY)

By: M. Berbers

Signature: