



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-2468
Expiration Date: 30-04-2026
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,41
Osmolality:	270 – 290 mOsm/kg	280
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	3,90
Gentamicin concentration:	5 – 15 mg/liter	9,74
MEA: (% blastocysts after 96 hours; 60 min. exposure time to test medium)	≥ 80%	100%
HSSA (motility compared with control after 24h exposure to test medium):	≥ 80%	100%

Approved on: 05-11-2024
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', with a circular stamp or mark above it.