



Certificate of Analysis

Product: SpermWash®

Batch number: SW-2536
Expiration Date: 31-12-2026
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,43
Osmolality:	270 – 290 mOsm/kg	281
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	3,90
Gentamicin concentration:	5 – 15 mg/liter	10,00
MEA:	≥ 80%	95%
(% blastocysts after 96 hours; 60 min. exposure time to test medium)		
HSSA (motility compared with control after 24h exposure to test medium):	≥ 80%	98%

Approved on: 23-06-2025
(DD/MM/YYYY)
By: M. Berbers

Signature: