



## Certificate of Analysis

### Product: SpermWash<sup>®</sup>

Batch number: SW-2602  
Expiration Date: 31-07-2027  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,45
Osmolality:	270 – 290 mOsm/kg	280
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	3,70
Gentamicin concentration:	5 – 15 mg/liter	13,00
MEA: (% blastocysts after 96 hours; 60 min. exposure time to test medium)	≥ 80%	90%
HSSA ( <i>motility compared with control after 24h exposure to test medium</i> ):	≥ 80%	91%

---

Approved on: 22-01-2026  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint rectangular stamp.