



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-309
Expiration Date: 31-08-2024
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,45
Osmolality:	270 – 290 mOsm/kg	279
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05
HSA concentration:	3,20 – 4,80 g/liter	3,60
MEA: (% blastocysts after 96 hours; 60 min. exposure time to test medium)	≥ 80%	100%

Approved on: 23-02-2023
(DD/MM/YYYY)
By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp or background.