



Certificate of Analysis

Product: SpermWash[®]

Batch number: SW-830
Expiration Date: 31-12-2019
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,43
Osmolality:	270 – 290 mOsm/kg	276
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	100%

Approved on: 09-07-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.