

**Certificate of Analysis
GM501 WASH**

Batchnumber	WASH-1714
Production date	22 December 2017
Expiry date	30 June 2018
Total volume (mL)	28500
Batchnumber HSA	B3AKA00593

	Results	Specifications
pH	7,32	7,20-7,50
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 09 January 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann