

Certificate of Analysis

Product name: 35mm Dish, Non-Treated

Product type: Oosafe® Plasticware IVD

Product code: OOPW-TF06

Batch/LOT No.: 09031

Product expiry date: 2029-10

| TESTING METHOD | ACCEPTANCE CRITERIA | RESULTS | CONCLUSION | DOCUMENT NO. |
|-----------------------------------|--|---------|-----------------------------|----------------|
| Mouse Embryo Assay Test (MEA) | Over 80% embryo development to the expanded blastocyst stage within 96 hours | Passed | Non-embryotoxic | SPAR-5130-1024 |
| Human Sperm Survival Assay (HSSA) | ≥75% sperm motility after 24 hours proven | Passed | Stable human sperm motility | S8289-1024SPAR |
| Limulus Amoebocyte Lysate (LAL) | <0.03 EU/device | Passed | Do not contain endotoxin | E4290-1024SPAR |

Quality Management system according to ISO 13485:2016

Sterilized by irradiation according to ISO 11137-1

Sterilization method: E-Beam

Sterility assurance level: 10⁻⁶

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 29.10.2024



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk
Corp. VAT no.: DK 30 89 85 75