

# Certificate of Analysis

**Product name: Oosafe® Disinfection Wipes**

**Product type: Biocide**

**Product code: OODW-70**

**Batch/LOT No.: 42304**

**Production date: 2024-05**

**Product expiry date: 2027-05**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid inside wipes	Clear liquid inside wipes	QC_4.1-0624
Odor / Olfactive	Odorless	odoless	PRD/F01
ph (2% active substance in water)/ ph meter	7,0 – 9,0	8,1	PRD/F01
Density, 20 °C, gr/cm <sup>3</sup> / Density measuring device	1,001 – 1,010	1,008	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4500-0624
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8057-0624SPAR

Quality Management system according to ISO 13485:2016

**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

**Issue date: 08.07.2024**



**Sparmed ApS**

Page 1 of 1