

Certificate of Analysis

Product name: Oosafe® Disinfectant Wipes

Product type: Biocide

Product code: OODW-70

Batch/LOT No.: 44106

Production date: 2024-10

Product expiry date: 2027-10

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid inside wipes	Clear liquid inside wipes	QC_4.1-1024
Odor / Olfactive	odorless	odorless	PRD/F01
ph (2% active substance in water)/ ph meter	7,0-9,0	7,7	PRD/F01
Density, 20 °C, gr/cm ³ / Density measuring device	1,001-1,010	1,009	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-5193-1124
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8315-1124SPAR

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 12.11.2024



Sparmed ApS