

Certificate of Analysis

Product name: Oosafe® Disinfectant Wipes

Product type: Biocide

Product code: OODW-70

Batch/LOT No.: 53105

Production date: 2025-08

Product expiry date: 2027-07

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid inside wipes	Clear liquid inside wipes	QC_4.1-0825
Odor / Olfactive	odorless	odorless	PRD/F01
ph (2% active substance in water)/ ph meter	7.0 - 9.0	7.8	PRD/F01
Density, 20 °C, gr/cm ³ / Density measuring device	1.001 – 1.010	1.007	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-1398-25-08
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SPAR-1399-25-08

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 22.09.2025



Sparmed ApS