

Certificate of Analysis

Product name: OOSAFE® CLASS IIA MEDICAL DEVICE DISINFECTANT

Product code: OOMD-01000

Batch/LOT No.: 60865.1-MD

Production date: 2019-05

Expiry date: 2022-05

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0-8,5	7,5	NA
Density, 20 °C, gr/cm ³ / DIN 51757	1,001-1,010	1,002	NA
Quaternary Ammonium Compounds / QUAT test stripe	2,4-2,8	2,6 mg/gr	NA
Sodium nitrite		≤ 0,1 mg/gr	NA
Additives		3,2 mg/gr	NA
Water		The rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-1761-0519	Passed	20-05-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S3845-0519SPAR	Passed	17-05-2019

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Qi Wei,
Quality Control Manager
SparMED ApS

Date:



Qi Wei