

Certificate of Analysis

Product name: OOSAFE® CLASS II MEDICAL DEVICE DISINFECTANT

Product code: OOMD-01000

Batch/LOT No.: 60875.1-MD

Production date: 2019-10

Expiry date: 2022-10

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0-8,5	7,4	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1,001-1,05	1,010	NA
Quaternary Ammonium Compounds / QUAT test stripe	2,4-2,8	2,6 mg/g	NA
Sodium nitrate		≤ 0,1 mg/g	NA
Additives		3,2 mg/g	NA
Water		the rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-2887-1119	Passed	15-11-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4133-1119SPAR	Passed	13-11-2019

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS

Date: 18-Nov-2019


