

Certificate of Analysis

Product name: Oosafe® Disinfectant for CO2 Incubators and Laminar Flow Hoods

Product code: OODIH

Batch/LOT No.: 60885

Production date: 2020-05

Expiry date: 2023-05

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0 – 8,5	7,8	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1,01 – 1,10	1,08	NA
Quaternary Ammonium Compounds / QUAT test stripe	2,4 – 2,8	2,6 mg/g	NA
Sodium nitrite		≤ 0,1 mg/g	NA
Additives		3,2 mg/g	NA
Water		The rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-4260-0720	Passed	17-07-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4536-0720SPAR	Passed	15-07-2020

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Florencia Garcia
Quality Control Department
SparMED ApS
Date: 20-07-2020


