

Certificate of Analysis

Product name: OOSAFE® DISINFECTANT FOR CO2 INCUBATORS AND LAMINAR FLOW HOODS

Product code: OODIH

Batch/LOT No.: 60915

Production date: 2021-11


Expiry date: 2024-11

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0 – 8.5	7.9	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1.01 – 1.10	1.09	NA
Quaternary Ammonium Compounds / QUAT test stripe	2.4 - 2.8	2,6 mg / gr	NA
Sodium nitrite		≤ 0.1 mg / gr	NA
Additives		3,2 mg / gr	NA
Water		The rest	
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-7514-0122	Passed	31/01/2022
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5611-0122SPAR	Passed	28/01/2022

Quality Control according to ISO 13458:2016

GOosafe with SparMED!
 Andrius Daugintis
 Quality Control Department
 SparMED ApS
 Date: 01/02/2022



 <p>GOosafe in IVF Laboratories</p>	PREPARED BY:	APPROVED BY:
	QUALITY CONTROL MANAGER	MANAGEMENT REPRESENTATIVE
	Signature <i>lvei</i>	Signature <i>Bmbe</i>