

Certificate of Analysis

Product name: OOSAFE® HAND DISINFECTANT

Product code: OODH

Batch/LOT No.: 80970

Production date: 2019-10

Expiry date: 2022-10

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	8,0-9,5	8,8	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1,001-1,05	1,030	NA
Quaternary Ammonium Compounds / QUAT test stripe	2,4-2,8	2,6 mg/g	NA
Additives		8,2 mg/g	NA
Water		the rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-2886-1119	Passed	10-11-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4134-1119SPAR	Passed	09-11-2019

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS
Date: 12-Nov-2019

