

## Certificate of Analysis

**Product name: Oosafe® Hand Disinfectant**

**Product code: OODH**

**Batch/LOT No.: 80980**

**Production date: 2020-05**

**Expiry date: 2023-05**

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid	Clear liquid	NA
Odor / Olfactive	Mild	Mild	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	8,0 – 9,5	8,7	NA
Density, 20 °C, gr/cm <sup>3</sup> / Density measuring device	1,01 – 1,10	1,07	NA
Quaternary Ammonium Compounds / QUAT test stripe	2,4 – 2,8	2,6 mg/g	NA
Additives		8,2 mg/g	NA
Water		The rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-4260-0720	Passed	17-07-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4536-0720SPAR	Passed	15-07-2020

Quality Control according to ISO 13458:2016

### GOosafe with SparMED!

Florencia Garcia  
Quality Control Department  
SparMED ApS  
Date: 20-07-2020


