

Certificate of Analysis

Product name: Oosafe® HAND DISINFECTANT

Product type: Biocide

Product code: OODH

Batch/LOT No.: 81050

Production date: 2023-11

Product expiry date: 2026-11

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0224
Odor / Olfactive	mild	mild	PRD/F01
ph (2% active substance in water)/ ph meter	8,0 – 9,5	9,3	PRD/F01
Density, 20 °C, gr/cm ³ / Density measuring device	1,01 – 1,10	1,08	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR3614-0224
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S7065-0224SPAR

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 22.02.2024



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK ☎ +45 39 40 25 03 ☎ info@sparmed.dk ☎ www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75