

Certificate of Analysis

Product name: Oosafe® Hand Disinfectant

Product type: Biocide

Product code: OODH

Batch/LOT No.: 81070

Production date: 11.2024

Product expiry date: 11.2027

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0325
Odor / Olfactive	mild	mild	PRD/F01
ph (2% active substance in water)/ ph meter	8.0-9.5	9.4	PRD/F01
Density, 20 °C, gr/cm ³ / Density measuring device	1.01-1.10	1.08	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-688-25-03
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SPAR-689-25-03

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 10.04.2025



Sparmed ApS

Page 1 of 1

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75