

Certificate of Analysis

Product name: Oosafe® Hand Disinfectant

Product type: Biocide

Product code: OODH

Batch/LOT No.: 81080

Production date: 2025-05

Product expiry date: 2028-04

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0725
Odor / Olfactive	mild	mild	F-08-3
ph (2% active substance in water)/ ph meter	8.0-9.5	9.4	F-08-3
Density, 20 °C, gr/cm ³ / Density measuring device	1.01-1.10	1.08	F-08-3

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-1249-25-07
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SPAR-1250-25-07

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jageniak

Quality Control Department

SparMED ApS

Issue date: 26.08.2025



Sparmed ApS

Page 1 of 1