

Certificate of Analysis

Product name: Oosafe Class IIA Medical Device Disinfectant Wipes

Product code: OOMDW-120

Batch/LOT No.: 90901-MD

Production date: 02-2019

Expiry date: 02-2022

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid inside wipes	Clear liquid inside wipes	NA
Odor / Olfactive	Odourless	Odourless	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0-9,0	8,5	NA
Density, 20 °C, gr/cm ³ / DIN 51757	1,001-1,010	1,004	NA
Quaternary Ammonium Compounds / QUAT test stripe	1,9-2,2 %	2 %	NA
Sodium nitrite	NA	NA	NA
Additives	NA	NA	NA
Water	NA	The rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-1533-0419	Passed	07-04-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S3781-0419SPAR	Passed	04-04-2019

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS



Inga

Date: 9 Apr 2019