

Certificate of Analysis

Product name: OOSAFE® DISINFECTION WIPES

Product code: OODW-70

Batch/LOT No.: 94506

Production date: 2019-11

Expiry date: 2022-11

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid, inside wipes	Clear liquid, inside wipes	NA
Odor / Olfactive	Odourless	Odourless	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0-9,0	8,5	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1,001-1,05	1,005	NA
Quaternary Ammonium Compounds / QUAT test stripe	1,9-2,2 %	2 %	NA
Water		the rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-3032-1219	Passed	15-12-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4186-1219SPAR	Passed	12-12-2019

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS
Date: 16-Dec-2019

