

Certificate of Analysis

Product name: Oosafe® Laboratory Surface Disinfectant

Product code: OODSF

Batch/LOT No.: 70964

Production date: 2023 - 05

Expiry date: 2026 - 05

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear, slightly yellow liquid	Clear, slightly yellow liquid	26.07.2023
Odor / Olfactive	Slight smell of propionic acid	Slight smell of propionic acid	26.07.2023
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	5.5 – 7.0	6.7	26.07.2023
Density, 20 °C, gr/cm ³ / Density measuring device	1.01 – 1.15	1.04	26.07.2023
Quaternary Ammonium Compounds / QUAT test			
Sodium nitrite		< 7.0 %	26.07.2023
Additives		< 6.0 %	26.07.2023
Water		The rest	N/A
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR – 2482 - 0823	Passed	22.08.2023
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S6686 – 0823SPAR	Passed	23.08.2023

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak


Quality Control Department

SparMED ApS

Date: 31.08.2023



Manuela Jagieniak

 GOosafe in IVF Laboratories	PREPARED BY:	APPROVED BY:
	QUALITY CONTROL MANAGER	MANAGEMENT REPRESENTATIVE
	Signature <i>lveii</i>	Signature <i>Bmbe</i>