

# Certificate of Analysis

**Product name: Oosafe® Medical Devices Disinfectant**

**Product type: Class IIa**

**Product code: OOMD**

**Batch/LOT No.: 60965.1-MD**

**Production date: 2024-05**

**Product expiry date: 2027-05**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0524
Odor / Olfactive	Mild	Mild	COA/F01
ph (2% active substance in water)/ ph meter	7,0 – 8,5	7,7	COA/F01
Density, 20 °C, gr/cm <sup>3</sup> / Density measuring device	1,01 – 1,10	1,08	COA/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4365-0624
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8009-0624SPAR

Quality Management system according to ISO 13485:2016


**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

**Issue date: 16.06.2024**



**Sparmed ApS**

Page 1 of 1

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75