

## Certificate of Analysis

### GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-1801PRG
Production date	25 January 2018
Expiry date	31 July 2018
Total volume (mL)	37100
Batchnumber HSA	B3AKA00593

	Results	Specifications
pH	7,29	7,20-7,50
Osmolality	277	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	09 February 2018
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*