

# Certificate of Analysis

**Product name:** Oosafe® Laboratory Surface Disinfectant

**Product type:** Biocide

**Product code:** OODSF

**Batch/LOT No.:** 70984

**Production date:** 2024-05

**Product expiry date:** 2027-05

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear, slightly yellow liquid	Clear, slightly yellow liquid	QC_4.1-0524
Odor / Olfactive	slight smell of propionic acid	slight smell of propionic acid	PRQ-F01
ph (2% active substance in water)/ ph meter	5,5 – 7,0	6,9	PRQ-F01
Density, 20 °C, gr/cm <sup>3</sup> / Density measuring device	1,01 – 1,15	1,04	PRQ-F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4366-0624
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8018-0624SPAR

Quality Management system according to ISO 13485:2016


**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

**Issue date:** 16.06.2024



**Sparmed ApS**

Page 1 of 1