

# Certificate of Analysis

**Product name: Oosafe® Medical Device Disinfectant**

**Product type: Class IIa**

**Product code: OOMD**

**Batch/LOT No.: 60975.1-MD**

**Production date: 11.2024**

**Product expiry date: 11.2027**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0425
Odor / Olfactive	mild	mild	COA/F01
ph (2% active substance in water)/ ph meter	7.0-8.5	7.7	COA/F01
Density, 20 °C, gr/cm <sup>3</sup> / Density measuring device	1.01-1.10	1.08	COA/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-(R1)-25-04
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SPAR-926-25-04

Quality Management system according to ISO 13485:2016

**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

**Issue date: 28.05.2025**



**Sparmed ApS**

Page 1 of 1

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75