

# Certificate of Analysis

**Product name:** Oosafe® Hand Disinfectant

**Product type:** Biocide

**Product code:** OODH-0500

**Batch/LOT No.:** 08081

**ID number:** 1855/D2

**Production date:** 2026-04

**Product expiry date:** 2028-03

| TESTING METHOD                                       | ACCEPTANCE CRITERIA | RESULTS      | DOCUMENT NO. |
|------------------------------------------------------|---------------------|--------------|--------------|
| Color / Visual                                       | Clear liquid        | Clear liquid | QC_4.1-0726  |
| Odor / Olfactive                                     | Mild                | Mild         | F-WI-08-01-4 |
| ph (2% active substance in water)/ ph meter          | 10.0-10.5           | 10.34        | F-WI-08-01-4 |
| Quaternary ammonium concentration / QUAT test stripe | 2%                  | 2%           | F-WI-08-01-4 |

| TESTING METHOD           | ACCEPTANCE CRITERIA                                                          | RESULTS | CONCLUSION                  | DOCUMENT NO.       |
|--------------------------|------------------------------------------------------------------------------|---------|-----------------------------|--------------------|
| Mouse Embryo Assay (MEA) | Over 80% embryo development to the expanded blastocyst stage within 96 hours | Passed  | Non-embryotoxic             | MEA.026.2615.2026  |
| Human Sperm Assay (HSA)  | ≥75% sperm motility after 24 hours proven                                    | Passed  | Stable human sperm motility | SSMA.026.2616.2026 |

Quality Management system according to ISO 13485:2016


**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

**Issue date:** 17.07.2026



**Sparmed ApS**

Page 1 of 1

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75