

Certificate of Analysis

Product name: Oosafe® Disinfectant

Product type: MD Class IIa

Product code: OOMD-01000

Batch/LOT No.: 58906.1.D5-MD

Production date: 2026-04

Product expiry date: 2029-03

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1-0726
Odor / Olfactive	Mild	Mild	F-WI-08-01-4
ph (2% active substance in water)/ ph meter	7.0-8.5	7.57	F-WI-08-01-4
Quaternary ammonium concentration / QUAT test stripe	2%	2%	F-WI-08-01-4

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	MEA.026.2518.2026
Human Sperm Assay (HSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SSMA.026.2519.2026

Quality Management system according to ISO 13485:2016


GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 16.07.2026



Sparmed ApS

Page 1 of 1

Rytmemarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75