

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1801
Production date	15 January 2018
Expiry date	31 July 2018
Total volume (mL)	1470

	Results	Specifications
pH	7,3	7,20-7,60
Osmolality	297	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	90	≥ 80 %

Release date	30 January 2018
--------------	-----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann