GYNEMED

Certificate of Analysis
GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-1803PRG
Production date	09 March 2018
Expiry date	30 September 2018
Total volume (mL)	42000
Batchnumber HSA	B3AKA00593

	Results	**Specifications**
pH	7,27	7,20-7,50
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	27 March 2018
Released by	Dr. Julia Heinzmann
Signature	p.p. Julia Heinzmann

GYNEMED GmbH Co. KG, Lübecker Str. 9, 23738 Lensahn, Germany
phone: +49 4363 90329-0, fax: +49 4363 90329-19, e-mail: info@gynemed.de