

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-1805
Production date	06 July 2018
Expiry date	31 January 2019
Total volume (mL)	31500

	Results	Specifications
pH	7,26	7,20-7,40
Osmolality	277	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Release date	23 July 2018
Released by	Benthe Brauer

Signature

