

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Intra-Uterine Insemination Cannula with mandrel #4220 IUI Plus

LOT number: 20180423**STERILIZATION date:** 28-04-2018**EXPIRY date:** 23-04-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 22.5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:
13-06-2018