

CERTIFICATE OF QUALITY

ITEM name and code	#4220 IUI Plus - Intra-Uterine Insemination Cannula with mandrel
LOT NUMBER	20180424
EXPIRY DATE	24 April 2022
STERILIZATION DATE	28 April 2018
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
13 June 2018