

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1808
Production date	31 August 2018
Expiry date	28 February 2019
Total volume (mL)	1490

	Results	Specifications
pH	7,37	7,00-7,60
Osmolality	419	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 18 September 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann