

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-1817
Production date	17 August 2018
Expiry date	31 August 2019
Total volume (kits)	600
Batchnumber HSA	K809B6671

	Results	Specifications
pH	7,53	7,30-7,60
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,111	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	03 September 2018
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann