

**Certificate of Analysis
GM501 SpermActive**

Batchnumber	WASH-1809PRG
Production date	03 September 2018
Expiry date	31 March 2019
Total volume (mL)	33700
Batchnumber HSA	B3AKA00593 / B3AKC0053

	Results	Specifications
pH	7,23	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 18 September 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann