

Certificate of Analysis

Product name: Oosafe® Hand Disinfectant

Product type: Biocide

Product code: OOHD

Batch/LOT No.: 81060

Production date: 2024-05

Product expiry date: 2027-05

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	DOCUMENT NO.
Color / Visual	Clear liquid	Clear liquid	QC_4.1- 0924
Odor / Olfactive	Mild	mild	PRD/F01
ph (2% active substance in water)/ ph meter	8,9 - 9,5	9,3	PRD/F01
Density, 20 °C, gr/cm ³ / Density measuring device	1.01 – 1,10	1, 08	PRD/F01

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-5025-0924
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8243-0924SPAR

Quality Management system according to ISO 13485:2016

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 02.10.2024



Sparmed ApS

Page 1 of 1