

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1901
Production date	07 January 2019
Expiry date	31 July 2019
Total volume (mL)	1000

	Results	Specifications
pH	7,34	7,00-7,60
Osmolality	422	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 22 January 2019

Released by Vanessa Bierlich

Signature