

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-1824
Production date	05 December 2018
Expiry date	31 December 2019
Total volume (kits)	443
Batchnumber HSA	B3AKC00053

	Results	Specifications
pH	7,49	7,30-7,60
Osmolality	277	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,194	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 02 January 2019

Released by Vanessa Bierlich

Signature