

**Certificate of Analysis  
GM501 PVP**

Batchnumber	PVP-1822
Production date	12 December 2018
Expiry date	30 September 2019
Total volume (kits)	830
Batchnumber HSA	B3AKC00053

	<b>Results</b>	<b>Specifications</b>
pH	7,37	7,20-7,60
Osmolality	302	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	229,2	> 170 cP
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	03 January 2019
Released by	Vanessa Bierlich
Signature	

P.P. 